


Surgical and Prosthetic Concept Bone Level Implant Z5-BL



Z-SYSTEMS
ceramic implants

Welcome to the world of ceramic implants

The Z-SYSTEMS implant system is the result of many years of clinical and laboratory experience since 2004. Safety is our foremost priority.

This basic information on the surgical and prosthetic procedure of the Z-SYSTEMS Implant System is intended to provide dentists, physicians, surgeons and dental technicians with a description of the most important surgical and prosthetic steps for the planning, treatment and procedure of the Z-SYSTEMS System. This manual cannot replace implantological and prosthetic training. It is assumed that the user is familiar with the implant procedure.

1	General information	
1.1	General aspects and important information	6
1.2	Material, biocompatibility and osseointegration	8
1.2	Indications	10
1.4	Fundamentals of treatment planning	12
1.5	Protective measures	14
2	Surgery	
2.1	Instruments	18
2.2	Sterilization	20
2.3	Surgical procedure/Drilling protocol	21
2.4	Specific features of Z5-BL	28
2.5	Implant removal	30
2.6	Post-operative Care	32
3	Prosthetic concept	
3.1	Fixing the abutment with Z5-BL implants	34
3.2	Impression with Z5-BL implants	37
3.3	Gingiva formers	40
3.4	Model fabrication	41
3.5	Temporary care of Z5-BL implants	42
3.6	Restoration with a laboratory-fabricated long-term temporary restoration after osseointegration	43
3.7	Final restorations on Z5-BL implants	44
3.8	Prosthetic restoration of Z5-BL implants	45
3.9	Prosthetic restoration of Z5-BL implants in the edentulous jaw	46
3.10	Prosthetic aftercare of the Z5-BL implants	49
3.11	Cementing of restorations on Z5-BL implants	50
3.12	Prophylaxis for Z5-BL implants	51
3.13	How to remove a ceramic screw fragment from a ceramic implant after a fracture	52

General aspects and important information

General features

Z-SYSTEMS implants are unique in their combination of design and material. We expressly ask you to read this manual before starting and to exactly follow our specific instructions on preparation as well as surgical and prosthetic procedures..

Observing these specific instructions and following the general implantological and prosthetic principles will ensure safe and successful implantation with Z-SYSTEMS implants.

The health of your patients is our top priority. For this reason we have compiled a technical guide that will contribute to the success of treatment with Z5-BL implants. The surgical and prosthetic phase should be preceded by extensive preoperative assessment, diagnosis and planning. Careful planning and adherence to the protocols for implantation and prosthetic restoration of Z5-BL implants reduces/ avoids problems/errors during implantation and especially during prosthetic restoration.

We recommend the use of Z5-BL implants only for dentists with thorough, practical and surgical training and with expertise and experience in implantology. Instruction / training by an implantologist or Z-SYSTEMS representative familiar with the use of the instruments is strongly recommended. Z-SYSTEMS offers regular training courses with experienced users for starting with the system.

Important information

Disclaimer: The Z5 implant system is part of an overall concept and may only be used in conjunction with the corresponding original components and instruments and according to the Z-SYSTEMS instructions and recommendations. Instructions regarding the application of our products are given verbally, in writing, electronically or

through practical training, in accordance with the state of the art at the time of product launch. The user of Z5-BL products must decide whether or not a product is suitable for a patient and a specific situation according to their indication. Z-SYSTEMS excludes any liability for damages resulting from the use or implantation of Z5-BL products as a result of, or in connection with, errors in professional assessment or application/indication, in particular also claims due to the disregard of general implantological and prosthetic principles in connection with implants. The user is also obliged to inform themselves regularly about the latest developments of our system and its applications.

Send us an e-mail to support@zsystems.com and we will be happy to send you the latest information.

Training: We recommend exchanging experiences, learning from and with colleagues. Z-SYSTEMS offer its users and interested parties extensive options for continuous professional development. Z5-BL courses are offered with and without live surgery.

Information regarding all our courses can be found at www.zsystems.com under «Events».

Availability: Not all of the products described in this manual are available in all countries. For further information, please contact our subsidiary or sales company in your country.

Precautions: Our products must be protected from aspiration during intraoral use.

Delivery: The sale of these products is limited to dentists, doctors or licensed dental technicians or orders made on their behalf.

Units per package: Unless otherwise stated, the package unit is 1 piece.

Documentation: Detailed instructions regarding the Z5-BL implant system are available from your account manager or customer service department in our headquarters.

Qualified users: Z-SYSTEMS implants should only be used by dentists, doctors, surgeons and dental technicians that are trained to use the system. Corresponding courses are offered by Z-SYSTEMS.

Certification: FDA/CE/ISO13485/MDD93/42 EWG














Z-SYSTEMS have complied fully since 2004 with the current normative and legal requirements for medical products through European certification according to ISO 13485, as well as the guideline 93/42/EEC for medical devices. Z-SYSTEMS have been registered with the FDA (US Food and Drug Administration) since 2007.

Colour coding of the surgical and prosthetic products:

- Yellow:** 3.6 mm diameter
- Red:** 4.0 mm diameter
- Green:** 5.0 mm diameter

Do not expose the products to direct sunlight.

Explanation of the symbols on labels and package inserts

	Batch number
	Item number
	Plasma sterilised
	Non-sterile
	Do not use if packaging is damaged
	Single use, not reusable
	Attention: Observe the package inserts
	Consult the instructions for use
	Use before expiration date
	Date of manufacture
	Manufacturer
	Z5-BL Products are CE marked and comply with the requirements of the Medical Device Directive 93/42 EEC.
	CAUTION: United States Federal Law restricts this device to sale to, or on the order of, a licensed dentist or physician.

Observe the package insert.

Material, biocompatibility and osseointegration

Material

All Z5-BL implants are manufactured according to the unique «Zirkolith®» process from zirconium oxide TZP-A bioceramics in compliance with the ISO 13356 standard – it encompasses our experience in the development, material processing, quality assurance and finishing of zirconium oxide. The composition and production processes for zirconium oxide vary according to the requirements for the system component, for example whether it is an implant, a cutting instrument or some other surgical instrument.

The material achieves its properties through the "Hot Iso-static Pressing" process. In this process, the material is recompressed in a tunnel kiln for three days at 2000 bar after the sintering process, which improves the physical properties of the base material.


Not only the implants, but also the instruments that come into direct contact with the bony surgical area are made of zirconium oxide. The cutting instruments are made of high-strength ATZ high-performance ceramics (Alumina Toughened Zirconia).

Biocompatibility



Numerous studies since the 1960s have confirmed the biocompatibility of zirconium oxide ceramics.

Osseointegration

Zirconium oxide has similar osseointegration behaviour to commercially pure titanium, which has also been proven in a large number of studies.



Hot Isostatic Pressing in a tunnel kiln:
Pressure up to 2.000 bar, temperatures up to 2.000 °C



Surface

Surface modification is made in the SLM® process developed by Z-SYSTEMS using laser technology and results in an increase in surface area and therefore to increased macro and micro roughness.

Healing time

We recommend a healing time of 3 months in the lower jaw and 6 months in the upper jaw* for healthy patients with good bone density and sufficient bone quality.

We strongly recommend that each implant is protected during the healing phase, for example through provisional arrangements or ground prostheses.

* Please note that all references to bone classification in this manual refer to the classification as described by Lekholm and Zarb: Lekholm U.R. & Zarb G.A.: Patient selection and preparation, in Brånemark P.-I., Zarb G.A., Albrektsson T. (eds): Tissue-Integrated Prostheses: Osseointegration in Clinical Dentistry. Chicago IL, Quintessence, 1985, PP 199–209.

Indications for Use

Z5-BL implants are designed for surgical implantation into the upper and lower jaw for the attachment of prosthodontic appliances to replace missing teeth. Z5-BL

implants are suitable for patients with metal allergies and the chronic diseases resulting from them. Z5-BL implants are intended for delayed loading.

General areas of application

As a rule of thumb, the implant with the largest possible diameter should always be used, because the mechanical strength increases disproportionately with increasing diameter of the implant.

Applications for 3.6 mm

Ø 3.6 mm implants are only approved for use in the lateral incisor region (tooth 7/10) of the upper jaw and in the incisor region (tooth 23/24/25/26) of the lower jaw. Their inclusion in bridge constructions are only permitted if each tooth to be replaced is with an implant and is located in the regions mentioned above.

The inserts are not suitable for applications where there is a risk of excessive cantilever like movement (e.g., single-tooth replacement for 8/9, molars, premolars, extended crowns, extension bridges, bridges, bar work, telescopic work).

Applications for 4.0 mm

Universal implant that is suitable for most indications. Not suitable for applications where there is a risk of excessive bending moments (e.g. extended crowns, extension bridges, bridges with more than one pontic). Limited suitability for extension bridges and telescopic restorations. Telescopic or Locator-type abutment restoration on at least 4 implants is recommended.

Applications for 5.0 mm

Universal implant, suitable for most indications where there is sufficient bone. Not suitable for applications where there is a risk of excessive bending moments (e.g. extended crowns, extension bridges, bridges with more than one pontic). Implants with Ø 5.0 mm are recommended for the indication canines, central upper incisors and upper jaw/lower jaw molars. Limited suitability for extension bridges and telescopic restorations. Telescopic or Locator-type abutment restoration on at least 4 implants is recommended.

Implant size	Thread diameter	Shoulder diameter	Minimum space requirements orovestibular (surgery)	Minimum space requirements mesio-distal (surgery)	Optimum indication odontogram	Single-tooth	Blocking	Bridge in premolar width (max. span 1 pontic)	Extension bridge	Bar	Telescope		
3.6	3.6 mm	3.6 mm	5.6 mm	5.6 mm	UPPER RIGHT 1st Quadrant	UPPER LEFT 2nd Quadrant		+	+	-	-	-	-
					USA	7	10						
					FDI	12	22						
					FDI	42 41	31 32						
					USA	26 25	24 23						
LOWER RIGHT 4th Quadrant		LOWER LEFT 3rd Quadrant											
4.0	4.0 mm	4.0 mm	6.0 mm	7.0 mm	UPPER RIGHT 1st Quadrant	UPPER LEFT 2nd Quadrant		+	+	+	-	+	(+)
					USA	2 3 4 5 6 7 8	9 10 11 12 13 14 15						
					FDI	17 16 15 14 13 12 11	21 22 23 24 25 26 27						
					FDI	47 46 45 44 43	33 34 35 36 37						
					USA	31 30 29 28 27	22 21 20 19 18						
LOWER RIGHT 4th Quadrant		LOWER LEFT 3rd Quadrant											
5.0	5.0 mm	5.0 mm	7.0 mm	8.0 mm	UPPER RIGHT 1st Quadrant	UPPER LEFT 2nd Quadrant		+	+	+	-	+	(+)
					USA	2 3 4 5 6 7 8	9 10 11 12 13 14 15						
					FDI	17 16 15 14 13 12 11	21 22 23 24 25 26 27						
					FDI	47 46 45 44 43	33 34 35 36 37						
					USA	31 30 29 28 27	22 21 20 19 18						
LOWER RIGHT 4th Quadrant		LOWER LEFT 3rd Quadrant											
+ recommended (+) not recommended - not possible													

Fundamentals of treatment planning

The patient must meet the generally valid implant surgery and prosthetic criteria for an implant restoration.

Implant prosthetic restoration is a collaboration involving the dentist/surgeon and dental technology and requires a high degree of clinical experience and detailed knowledge from all involved.

The following are important planning points:

Z-SYSTEMS recommend the selection of the appropriate implant and its restoration according to the following criteria:

- Endosseous diameter of the implant
- Shoulder diameter of the implant
- Length of the implant
- Vertical implant position

Aesthetically optimum result


Many conditions are decisive for an aesthetically optimum result:

- the harmonious course (contour) of the gingiva
- the best implant position (vertical, orofacial and mesio-distal)
- the shape of the crown and
- the presence of interdental papillae

Planning the position of the implant

During planning, the instructions for the hard tissue configurations are to be complied with and soft tissue management must be observed.

The implant diameter and implant length must be determined so that there is sufficient bone (at least 1mm) around the implant. A minimum distance of 1.5mm to an adjacent natural tooth and 3mm to an adjacent implant must be maintained.



Structure-preserving and structure-protecting procedures are to be used for flap design and implant placement. The oral hygiene requirements must be taken into account as early as the planning stage.

Restorations

Single-tooth crowns


Restoration with single crowns is a possible restoration under the aspect of "restitutio ad integrum" ("restoration to original condition"). It includes all the advantages that are possible in periprosthetic rehabilitation.

The physiologically adequate biomechanical load prevents further atrophy of the hard and soft tissue.




Splinted crowns

Splinting of the crowns may be necessary for static reasons (such as unfavourable lever ratios). When selecting splinting, maintenance of good hygiene must be considered.



Implant-supported bridges


Implant-supported bridges can be inserted in positions that do not permit implant placement. The implant distribution must be selected so that small span segments are created.



Locator-type abutment restoration

Locator-type abutments are for attachment of overdenture prostheses into the edentulous upper and lower jaw.

- The principle of 4-point-support is recommended
- Avoidance of axis divergence



Planning and specific features of applications with locator-type abutments

This page must be followed for the treatment of patients with locator-type restorations.

NOTE: To ensure optimal performance of the retentions and avoid loading the implants beyond their stability, strive for an axial transfer of force to the implants. For this purpose, the implants should be positioned as parallel to each other and perpendicular to the occlusal plane as possible. If practical, the implants should be placed in the same horizontal plane to allow easy handling when removing or inserting the prosthesis.

Guided surgery

When case planning, Z-SYSTEMS recommends using 3-dimensional x-ray images (CB/CT) and referencing the drill guide. Planning in this way will allow for a more axial alignment and assist with parallelism.

Gingiva height

Before surgery, measure the maximum tissue thickness at the planned implantation site (e.g., using a probe and attached measuring stop root canal instrument, local anesthesia).


Implant divergence

After pilot drilling with the DP230 depth gauge, Z-SYSTEMS recommends a visual check of the axis alignment for parallelism.

The maximum divergence between multiple implants is 40° and the maximum angulation per implant body is 20°. For implants angulated or near the maximum recommended angulation, we recommend the use of the Angle LC abutments which have an angulation of 15°. If there is a divergence of more than 20° per implant to the occlusal plane,


or more than 40° between several implants, the axial alignment must be corrected.

To ensure optimal performance of the Novaloc®-matrix, the working area is 1.35 mm above the surrounding gingiva (1.85 mm if the overdenture is to be made with 0.5 mm gingival clearance).



Avoidance of axis divergence

There are straight and angled locator-type abutments in two different gingival heights (GH).



Protective measures

For successful osseointegration, the implants must be protected from macro movements during the healing phase. Depending on bone quality, insertion torque, periostest measurement and general patient compliance, the dentist decides whether and which additional protective measures are necessary. Possible protective measures are: protective splints, splinted temporary restorations or protective prostheses.

Examples of protective measures are available on request by email at support@zsystems.com.

2 Surgery

2.1	Instruments	18
2.2	Sterilization	20
2.3	Surgical procedure/Drilling protocol	21
2.4	Specific features of Z5-BL	28
2.5	Implant removal	30
2.6	Post-operative Care	32


Instruments

The Z5-BL surgical instruments from Z-SYSTEMS are to be used.


The instruments required for implantation and have been designed to be user-friendly. The rotating instruments are marked with a colour code.

The instruments are labelled with the respective instrument designation to avoid any risk of confusion.


Driver




TP-SN/TP-SNS
Transfer-piece non-sterile/sterile with snap in mounting




SD-BS-S
Single-use screwdriver short, for basal screw



SD-BS-L
Single-use screwdriver long, for basal screw



BL-SD-ST
Screwdriver short for internal hex head




BL-SD-LT
Screwdriver long for internal hex head

Gauge



DP230
Depth gauge 2.30 mm



DP285
Depth gauge 2.85 mm



DP325
Depth gauge 3.25 mm



DP375
Depth gauge 3.75 mm




DP425
Depth gauge 4.25 mm

Meaning of the colours:
yellow = ø 3.6 mm, red = ø 4 mm, green = ø 5 mm


Material properties

All instruments that come into direct contact with the surgical field are made of zirconium oxide. The cutting instruments are made of high-strength ATZ high-performance ceramic (Alumina Toughened Zirconia). This alumina-reinforced zirconium oxide is ideal for the manufacture of drills and taps. The ATZ drills cut excellently with very little wear. Note: The drills must be replaced after being used 20 times.


Drills




RD230
Round burr 2.3 x 16 mm




TD230
Twist drill 2.3 x 16 mm




TD285
Twist drill 2.85 x 16 mm



TD325
Twist drill 3.25 x 16 mm



TD375
Twist drill 3.75 x 16 mm



TD425
Twist drill 4.25 x 16 mm

Cortical drill



CD360
Cortical drill 3.6 mm



CD400
Cortical drill 4 mm



CD500
Cortical drill 5 mm

Tap



T-360-3
Tap 3.6 mm



T-400-3
Tap 4 mm



T-500-3
Tap 5 mm

Adapter for tap



ZT-HA-9
Contra-angle adapter



ZT-RA10-9
Ratchet adapter 10 mm



ZT-RA20-9
Ratchet adapter 20 mm

Ratchet



TR-70
Torque ratchet

Accessories



TD-DS230
Drill stop 2.3 mm



TD-DS285
Drill stop 2.85 mm



TD-DS325
Drill stop 3.25 mm



TD-DS375
Drill stop 3.60 mm 3.75 mm



TD-DS425
Drill stop 4 mm 4.25 mm



CD-DS500
Drill stop 5 mm



ZT-HE14 (KI589)
Contra-angle extension



BL-CD
Loosening instrument

Sterilization

Sterilisation in a steam steriliser/autoclave

Use steam sterilisation processes with a fractionated vacuum process (and sufficient product drying). Other sterilisation methods (including gravitational steam sterilisation) are not permitted. Pay attention that:

- the sterilisation temperature does not exceed 138°C/280°F (plus tolerance in compliance with DIN EN ISO 17665).
- EU: the sterilisation holding time (exposure time at sterilisation temperature) is at least 4 minutes at a minimum temperature of 134°C/273°F.
- USA: the exposure time is 4 minutes at 132°C/270°F.
- All zirconia abutments, abutment screws, gingiva formers and temporary abutments are to be wrapped in a sterilization wrap that is FDA-cleared for the indicated cycle and dried for 30 minutes.

Surgical procedure/Drilling protocol


General drilling protocol

General note:


Round burr
To predrill the cortical bone and fix the implant position.

Twist drill
The implant bed is prepared with the twist drills in ascending order. The last drill used depends on the diameter of the implant to be inserted. Please follow the detailed instructions. The depth markings on the drill are easy to read. The first depth mark is 8 mm.

CAVE: The apical excess length of the drill tip is maximum 0.8 mm longer than the insertion depth of the implant. Please take this into account during the drilling process.



Drill stops
Drill stops are available for twist drills and cortical drills in the respective diameters. These are attached to the corresponding drills from the contra-angle handpiece connection side in the direction of the arrow and fixed at the required drilling depth. To remove, simply pull off in the direction of the arrow.



Cortical drill
Cortical drills are available to expand the cortical area according to the implant diameter. The use of a cortical drill is recommended for cases with hard bone or hard cortical bone.

Tap
In principle all Z5-BL implants are self-tapping. The use of a tap is recommended for cases with hard bone or hard cortical bone.

The general recommendations are:

Bone class D1+D2:

- Cortical area expansion with the cortical drill up to the depth marking
- Tap the entire length


Bone class D3+D4: do not tap

Exemplary procedure: Preparing the implant bed

The following shows how to prepare the implant bed using the example of a Z5-BL implant \varnothing 4.0mm/10mm in hard bone (D1).


After unfolding the gingiva, the basic preparation of the implant bed begins with preparation of the alveolar ridge and marking the implantation site with a round burr (RD230). This is followed by the pilot drill with the twist drill (TD230) and the further preparation of the implant bed using the twist drills in accordance with the endosteal implant diameter.

The threads are pre-cut with the tap; please refer to the notes on the previous page.


1. Preparation of the alveolar ridge and marking of the implantation site

Carefully reduce and smooth a narrow and tapered alveolar ridge with the RD230 round burr. This results in obtaining a flat and sufficiently wide bone surface. Mark the implantation site determined during planning of the implant position with the RD230 round burr.


Note: Depending on the clinical situation, this step may be omitted or applied in a modified form (e.g. for fresh extraction sockets).

2. Implant axis and depth


Use the twist drill TD230 to mark the implant axis by drilling to a depth of about 5mm. Use the depth gauge DP230 to check the correct orientation of the implant axis. Drill the implant bed to the final preparation depth with the twist drill TD230. If necessary, correct the orientation of the implant axis.




Use depth gauge DP230 to check the implant axis and preparation depth. Take an x-ray at this time, especially if the vertical bone volume is reduced. The depth gauge is inserted into the drilled hole and allows a visual assessment of the hole in relation to the anatomical structures.

3. Widening the implant bed to \varnothing 2.85 mm


Widen the implant bed with twist drill TD285.

4. Widening the implant bed to \varnothing 3.25 mm

Widen the implant bed with twist drill TD325.


5. Profile drilling

The cortex is widened to the diameter of the implant with the CD400 cortical drill.


6. Tap

Pre-tap the thread with the T400-3 tap over the entire length of the implant bed preparation for bone class D1+D2.


Drilling protocol



Drilling depth according
to the implant length:
8/10/12 mm



Drills with depth stops



Specific features of Z5-BL

Concept

The threaded Z5-BL implant is a self-tapping bone level implant. The Z5-BL implant has no transgingival portion (shoulder) and is surgically placed at bone level. Inside the implant there is a thread in which components such as healing caps, gingiva formers and abutments are fixed with the aid of an basal screw. After implantation, the inner lumen of the Z5-BL implant is closed with the supplied healing cap (BL-HC) made of radiopaque PEEK (polyether ether ketone) by simply screwing in the healing cap for submerged healing. A selection of standard gingiva formers is available for shaping the soft tissue before the prosthetic restoration. An individual design of the emergence profile can be achieved with the help of the temporary abutment and a temporary crown. Straight and angled standard abutments, the Z-abutment, crown and bridge abutments as well as Locator-type abutments are available for the final prosthetic restoration.

During surgery the clinician decides the dimensions of the preparation, depending on the bone quality. The recommended drilling protocol must be observed.


The optimum insertion torque is in the range of 25–35 Ncm. For harder bone, a tap should be used to avoid torques of over 35 Ncm when inserting. The twist drills have a depth stop to ensure safe and precise preparation of the implant osteotomy.

Implant removal from the sterile packaging


After opening the secondary packaging, removing the sterile inner blister and open the sealed lid. The white implant holder is rotated clockwise, and the implant is now easily accessible.

The implant driver is a two-piece component. The transfer-piece (TP-SNS) is inserted into the preferred adapter (ZT-HA-9, ZT-RA10-9, ZT-RA20-9) until it clicks into place.

Firmly press the corresponding adapter with the TP-SNS into the implant, taking into account the hexagon. Now the implant can be removed and inserted into the prepared osteotomy. After insertion, the implant driver must be removed again.



The alignment of the angled abutment can be defined with the transfer-piece. The hexagon on the ratchet adapter and transfer-piece correspond to the six sides of the internal geometry, as highlighted. The angled abutments overlap the hexagonal surface.



Healing phase

After implantation, the inner lumen of the Z5-BL implant is closed with the supplied healing cap (BL-HC) made of PEEK (polyether ether ketone) with a simple, hand-tight screw in using the screwdriver. Whenever possible, a submerged healing is desired. A good wound closure with tightly fitting gingiva is important. No forces should be exerted on the Z5-BL implants during the submerged healing phase. No other special protective measures are necessary. After the healing phase, the implants are exposed and the healing cap removed.



Implants

A total of nine different Z5-BL implants are available. Three diameters, 3.6, 4.0 and 5.0 mm, each in lengths of 8, 10 and 12 mm.

3.6 mm diameter Z5-BL-3608 Bone level implant Ø 3.6 mm, 8 mm long	 Z5-BL-3610 Bone level implant Ø 3.6 mm, 10 mm long	 Z5-BL-3612 Bone level implant Ø 3.6 mm, 12 mm long
4 mm diameter Z5-BL-4008 Bone level implant Ø 4 mm, 8 mm long	 Z5-BL-4010 Bone level implant Ø 4 mm, 10 mm long	 Z5-BL-4012 Bone level implant Ø 4 mm, 12 mm long
5 mm diameter Z5-BL-5008 Bone level implant Ø 5 mm, 8 mm long	 Z5-BL-5010 Bone level implant Ø 5 mm, 10 mm long	 Z5-BL-5012 Bone level implant Ø 5 mm, 12 mm long

Removing the implant from the packaging

All Z-SYSTEMS implants are delivered in a sturdy cardboard box. Inside is an outer blister (secondary packaging), with the inner blister (primary packaging) and the implant, as well as the package insert and three removable label strips for documentation.

1

TP-SN/SNS Adapter

2

Carefully apply light pressure to stabilize the white insert.

3

Open the sealed lid

4

Turn the implant holder counterclockwise

5

Slide the transfer piece into the implant with a slight rotational movement. A click will be heard when the transfer piece is properly attached.


6

For safe removal, ensure that there is no gap between the TP-SNS and the implant shoulder

7

Remove the implant from the insert by hand or by attaching the ratchet or handpiece

Implant placement




Transfer-piece TP-SNS driven by:
ZT-HA-9 | ZT-RA10-9 | ZT-RA20-9

Insertion speed (rpm)	15
max. torque	25-35


USER TIP:
Turn the implant slightly to the left before insertion. The thread noticeably engages in the osteotomy and then follows the threads in a clockwise direction as it is inserted.

Implant Driver Removal



After the implant placement, remove the transfer piece from the implant with the adapter still mounted completely.

Removal of a broken Transfer-piece



The transfer-piece is provided with a predetermined breaking point to prevent excessive tightening torque. If the transfer-piece breaks during insertion, one fragment remains in the implant and one in the adapter. To extract the implant, simply take out the broken transfer-piece part from the adapter with tweezers, re-insert the adaptor on the transfer-piece part in the implant. Counterclockwise turns will remove the implant. To ensure an insertion torque of 35 Ncm, the implant bed preparation must be checked and re-prepared, to avoid bone overcompression.

Transfer-piece for Z5 BL/TL Implants

Retentive notch

• to ensure secure retention to adapter i.e. handpiece or ratchet

Predefined breaking point

• to ensure excess torque is not applied to the implant.

Abutment alignment indicator

• Alignment aid for ideal prosthetic abutment orientation
Congruent hexagon (Implant – TP-SNS – Abutment)


Retentive notch

• to ensure secure retention to adapter i.e. handpiece or ratchet after failure of the predefined breaking point

Snap feature/Retention

• to ensure secure retention of the implant
• Detaches with Adapter after implant insertion.

Implant extraction after breaking the pre-defined breaking point to check the implant bed preparation



Post-operative Care

Postoperative care protocol

The following postoperative checks should be carried out at the intervals indicated:

Regular hygiene examinations (depending on the oral hygiene of the patient) up to the beginning of the prosthetic restoration.

Consultation with the surgeon to determine the return visit schedule during the first 6–8 weeks of the healing phase. Depending on the case, further conditioning of the soft tissue can be performed with the aid of a gingiva former before the final impression is taken.

The patient should be instructed to contact the practice immediately in the event of any complaints. A prophylactic check should be carried out 14 days and 6 weeks after implantation, and at 3 months at the latest.

Successful integration:

- No peri-implantitis
- No clinically noticeable loosening of the implant
- No pain in the vicinity of the implant
- No radiographic visible peri-implant gap

3 Prosthetic concept

3.1	Fixing the abutment with Z5-BL implants	34
3.2	Impression with Z5-BL implants	37
3.3	Gingiva formers	40
3.4	Model fabrication	41
3.5	Temporary care of Z5-BL implants	42
3.6	Restoration with a laboratory-fabricated long-term temporary restoration after osseointegration	43
3.7	Final restorations on Z5-BL implants	44
3.8	Prosthetic restoration of Z5-BL implants	45
3.9	Prosthetic restoration of Z5-BL implants in the edentulous jaw	46
3.10	Prosthetic aftercare of the Z5-BL implants	49
3.11	Cementing of restorations on Z5-BL implants	50
3.12	Prophylaxis for Z5-BL implants	51
3.13	How to remove a ceramic screw fragment from a ceramic implant after a fracture	52


Fixing the abutment with Z5-BL implants

Abutments

The abutments should be selected by the dentist, taking into account the previous prosthetic planning. The implant axis, the gingival height and the occlusion concept must be taken into account.

The following abutments are available:

- Straight standard abutments in two different gingiva heights for cementable single-tooth crown and multiunit bridge restorations.



- 15° angled abutments in two different gingiva heights and an anatomically adapted shape for cementable single-tooth crown and multi-unit bridge restorations.



- Crown and Z abutments in three different widths and two different gingival heights for cementable single crown and bridge restorations. The scanbody (BL-SB-36) can be visually detected with suitable dental scanners and subsequently used in the modelling and fabrication of final prostheses (direct crowns) using CAD/CAM techniques.




Explanation of abbreviations:

- ∠ Angle dimension
- Ø Platform diameter
- GH Gingiva height
- AH Abutment height
- CA Cone angle

Meaning of the colours:

yellow = Ø 3.6 mm, red = Ø 4 mm, green = Ø 5 mm



- Bridge abutments in two different gingiva heights without indexing (anti-rotation) in the implant-abutment connection for cementing bridges.



- Straight and angled Locator-type abutments in two different gingiva heights for anchoring of removable implant-supported full dentures in the edentulous jaw.



All zirconia abutments and basal (abutment) screws are supplied non-sterile and must be sterilised before use in the patient.



Indication restriction:

The abutments BL-ZB1538 and BL-AN1515 are exclusively for the Ø 3.6 mm bone level implant.

Fixing the abutments

Firmly press the abutment into the implant body by hand. Make sure that the abutment engages in the hexagon.

The connection between abutment and implant is secured by screwing a basal screw into the internal thread of the implant body. A short as well as a long screwdriver are available for the basal screws (SD-BS-S and SD-BS-L). The maximum permissible tightening torque value is reached when the handle of the screwdriver is turned off.

The thread is reversible and can be loosened again. In addition, if necessary, the cone disconnect instrument (BL-CD) may also be used to remove the abutment from the implant.



To prevent accidental swallowing of the screwdriver pin or its handle, both parts should be secured with a dental floss loop.


Loosening the abutment

Since the conical implant-abutment connection has a very high accuracy of fit, there is a positive fit between implant and abutment. To be able to loosen the abutment safely again, the loosening instrument (BL-CD) must be used.



Basal screw

Two different types of screws are available. A ceramic basal screw (BL-OSC-H) or a gold anodized titanium basal screw (BL-OST-H) is available. In order to reach the correct torque value, both screw types, ceramic and titanium, must be tightened until the handle of the disposable screwdriver turns off.



The provisional abutment BL-TA0030 or the diameter reduced BL-TA1538 can be used until final restoration, by using the Basal screw BL-OST and the screwdriver BL-SD-ST or BL-SD-LT – with a maximum tightening torque of 10–15 Ncm.





Impression with the Z5-BL implant

Precise and rotationally stable transfer parts are available for impression taking at implant level. An impression post for the closed impression (BL-IP-C), one for the open impression (BL-IP-O), and a scan body for the digital impression (BL-SB-36).

Note BL-SB:
Z-SYSTEMS cannot currently guarantee the provision of individual abutments. Z-SYSTEMS does not assume any guarantee for externally manufactured parts on original Z-SYSTEMS parts.
























The laboratory analogue BL-L is available for a conventional reconstruction on the stone model and a repositionable analog BL-L-3D for the digital work flow and 3D print.



Gingiva formers

A selection of gingiva formers with different gingiva heights and widths is available for shaping the soft tissue before prosthetic restoration. These are screwed into the implant hand-tight using a screwdriver. An individual design of the emergence profile can be achieved with the help of the temporary abutment and a temporary single-tooth restoration. All gingiva formers are supplied non-sterile and must be sterilised before use in the patient.


Gingiva former-assortment

 BL-GF1538 Gingiva former, GH 1.5 mm, Ø 3.8 mm	 BL-GF3055 Gingiva former, GH 3.0 mm, Ø 5.5 mm	 BL-GF1545 Gingiva former, GH 1.5 mm, Ø 4.5 mm	 BL-GF2545 Gingiva former, GH 2.5 mm, Ø 4.5 mm	 BL-GF1555 Gingiva former, GH 1.5 mm, Ø 5.5 mm	 BL-GF2555 Gingiva former, GH 2.5 mm, Ø 5.5 mm	 BL-GF2565 Gingiva former, GH 2.5 mm, Ø 6.5 mm
	 BL-A0015 Straight abutment, GH 1.5 mm, Ø 5.0 mm	 BL-BB1545 Bridge-Abutment, GH 1.5 mm, Ø 4.5 mm	 BL-BB2545 Bridge-Abutment, GH 2.5 mm, Ø 4.5 mm	 BL-CB1555 Crown-Abutment, GH 1.5 mm, Ø 5.5 mm	 BL-CB2555 Crown-Abutment, GH 2.5 mm, Ø 5.5 mm	 BL-CB2565 Crown-Abutment, GH 2.5 mm, Ø 6.5 mm
	 BL-A0025 Straight abutment, GH 2.5 mm, Ø 5.5 mm	 BL-CB1545 Crown-Abutment, GH 1.5 mm, Ø 4.5 mm	 BL-CB2545 Crown-Abutment, GH 2.5 mm, Ø 4.5 mm	 BL-TA0030 Temporary abutment, GH 3.0 mm, Ø 5.5 mm (with BL-OST screw) BL-TA1538 Temporary abutment, GH 1.5 mm, Ø 3.8 mm (with BL-OST screw)		
 BL-ZB1538 Z-Abutment, GH 1.5 mm, Ø 3.8 mm	 BL-A1515 15° angled abutment, GH 1.5 mm, Ø 5.0 mm	 BL-ZB1545 Z-Abutment, GH 1.5 mm, Ø 4.5 mm	 BL-ZB2545 Z-Abutment, GH 2.5 mm, Ø 4.5 mm			
 BL-AN1515 15° angled abutment, GH 1.5 mm, Ø 4.3 mm	 BL-A1525 15° angled abutment, GH 2.5 mm, Ø 5.5 mm					

Model fabrication

Fabrication of the master model
The matching laboratory analogue is inserted into the impression cap so that the lab analogue clicks into the impression cap with a perceptible click. This is the only way to ensure that the situation in the mouth is correctly represented in the master model.

The impression can then be cast with plaster and the master model can be completed.



Inserting the laboratory analogue into the impression before casting with plaster.

Temporary care of Z5-BL implants

General note

The general information on implant-supported restoration applies to the temporary restoration of Z5-BL implants.

Occlusion contacts must always be set so that a simple shim-stock foil can be pulled through interocclusally with slight resistance in the final bite position with maximum intercuspidation. Occlusion contacts should be point-shaped. Flat contacts must be avoided. A group function must be aimed for to relieve a single implant in the canine position.

If temporary restorations are to remain in place for a longer period of time, a close inspection of the firm hold and the static and dynamic occlusion and the periodontal conditions with any appropriate corrections and prophylactic sessions must be ensured. Temporary restorations on Z5-BL implants must have a passive fit.

Direct temporary restoration

Two different procedures are recommended for the fabrication of direct temporaries on Z5-BL implants in the mouth:

- Fabrication of a temporary restoration using an anatomic impression taken directly in the mouth
- Restoration with egg shell temporary

Restoration with a laboratory-fabricated long-term temporary restoration after osseointegration

If a temporary restoration on Z5-BL implants is intended to stay in place for a longer period (several months), it is recommended to use laboratory-fabricated, framework-reinforced long-term temporaries for stability reasons. The laboratory requires precise impressions for their fabrication.

The long-term temporary restoration must be completely stress-free and must have sufficient space for the placement of cement. Occlusion and dynamic occlusion must be precisely adjusted.

Procedure

- Check the passive fit of the long-term temporary restoration
- Check the aesthetics, form, phonetics
- Check the occlusion and dynamic occlusion
- Cement

Final restorations on Z5-BL implants

General note

Z5-BL implants can be restored with all restorative materials used in modern dentistry.

In addition to all-ceramic restorations, composites, metal restorations and combinations (PFM) are also conceivable. All restoration types are permanently cemented in the conventional manner.

Adhesive cementation of restorations to Z5-BL abutments is not possible. When restoring Z5-BL implants, the generally applicable guidelines for the planning and fabrication of implant-supported prosthetics must be observed.

The static, occlusal contact of the restoration must be considered in relation to the contacts of permanent teeth. The movement of permanent teeth must be taken into account, particularly with single-tooth restorations. Dynamic occlusal contacts on the restoration must be avoided. A group function must be aimed for to relieve a single implant in the canine position. A sufficient number of the supporting abutments and a statically favourable distribution must be ensured, as well as good cleaning possibilities.

Indication for the final prosthetic restoration of Z5-BL implants

The following clinical or radiographic findings indicate that the final prosthetic restorations can be fitted:

- No peri-implantitis
- No clinically noticeable loosening of the implant
- No loosening when attempting to unscrew (max. 15 Ncm/anaesthesia)
- No pain in the vicinity of the implant
- No radiographic visible peri-implant gap

Prosthetic restoration of Z5-BL implants

The valid general guidelines for the fabrication of fixed restorations on implants must be observed on Z5-BL implants. This applies in particular to the static and dynamic occlusion and the periodontium-prophylactic design of the restoration.

Indication of single-tooth restoration on Z5-BL implants

Z5-BL implants allow a restoration with fixed single-tooth crowns in the anterior and posterior regions.

The indication guidelines for implant selection must be observed. Furthermore, the instructions for restorations on Z5-BL implants with regard to static and dynamic occlusion, the periodontium-prophylactic design of the restoration, as well as the valid general guidelines for the fabrication of fixed restorations on implants must be observed.


Restoration of interdental gaps on Z5-BL implants

Fixed restorations can be placed on Z5-BL implants to close interdental gaps. Please note the preoperative selection of Z5-BL implants according to the Z-SYSTEMS indication guidelines and the sufficient number of abutments according to generally applicable prosthetic guidelines.

The mesial and/or distal extension of the restoration is not permitted under any circumstances. The integration of Z5-BL implants in composite bridges requires the exact observance of current dental implant clinical practice recommendations.



Single-tooth restoration of a front tooth with a Z5-BL implant




Prosthetic restoration of Z5-BL implants in the edentulous jaw

Restoration of Z5-BL implants with a bar construction

When planning a prosthetic restoration of Z5-BL implants using bar construction and removable prosthesis, the indication guidelines for implant selection must be observed. Number and location of implants (Z-SYSTEMS recommends 6 implants in the upper jaw, 4 implants in the lower jaw, min. 5 mm diameter) and the design of the prosthesis body and occlusion should depend on anatomical, functional and hygienic aspects.

The task of a bar restoration

- Stabilisation and primary blocking of the implants
- Securing the prosthesis against pulling and levering forces
- Force distribution
- Resilience compensation through degrees of freedom



Schematic diagram: No inclined arrangement of the bar link

The relining of an implant-supported bar prosthesis

Hybrid prostheses with resilient anchoring elements must be checked in a recall examination approximately every three months, to remedy any damaging movement of the prosthesis at an early stage using appropriate measures (such as relining).

Restoration of Z5-BL implants with a telescopic construction

In principle, the Z5-BL implants can be restored with telescopic constructions in combination with removable prostheses and bridges. However, there is an increased risk of forces not applied through the axis (especially high shear forces) acting on the implants. The abutments must be distributed so that at least one telescope is located at the distal end of the prosthesis (masticatory centre) so that no resiliencies act on the implants. A minimum implant diameter of 4 mm and a minimum number of 4 implants must be complied with. The integration of Z5-BL implants in telescopic construction requires the exact observance of current dental implant clinical practice recommendations.




Schematic diagram

Prosthetic restoration of Z5-BL implants with locator-type abutments

The new fabrication of a prosthesis is always recommended as part of overall planning or after implant restoration.

When fabricating the overdenture, please follow the detailed instructions of the manufacturer Valoc (www.valoc.ch) for the assembly of Novaloc™ matrices.



Schematic diagram: Z5-BL implant with BL-LC0030 abutment, Novaloc™
Section through matrix and prosthesis

For chair-side matrix fixation into the denture, prevent any resin from entering in between the matrix and the implant abutment. This can be accomplished by placing thin foil or a rubber dam between abutment and matrix. Make sure to provide sufficient space for the matrix and the resin in the overdenture.

Try-in of the overdenture

The try-in should take place at first without the retention inserts installed in the Novaloc™ matrices. In the first step you should check the fit of the overdenture on the gingiva and in occlusion. In the second step the denture is tried in with built-in retention inserts and the retention force is adjusted.

Matrix

For prosthetic restoration of Z5-BL implants with locator-type abutments only original Novaloc™ matrices of the manufacturer Valoc (www.valoc.ch) are recommended.



Novaloc™ PEEK matrix,
manufacturer: Valoc (www.valoc.ch)

Matrix housings are available from PEEK or titanium.

Colour-coded retention inserts are available with different retention values (red extra light/white light/yellow medium/green strong/blue extra-strong). The retention value (pull-off strength) can be varied easily by simple exchange of the retention insert. Please follow Valoc's (www.valoc.ch) manufacturer's instructions.

Impression taking

The impression can be taken with the Z5-BL impression components or with the Novaloc™ impression cap. A Novaloc™ forming/fixing matrix (impression cap) is available for impression taking. Please follow the relevant instructions of the manufacturer Valoc (www.valoc.ch).


Impression taking by Z5-BL impression components

- Impression with bone level impression components
- Master model production with the lab analog (BL-L)
- Fix the locator-type abutment with the laboratory screw into the lab analog (BL-L)
- place the Novaloc™ processing spacer* on the locator-type abutment
- fabrication of the prosthetic restoration

Impression taking by Novaloc™ forming/fixing matrix*

- Fix the locator-type abutment with occlusal screw in the implant
- Put on the Novaloc™ forming/fixing matrix*
- take impression
- Master model production with straight or angled Novaloc™ model analogue*
- place the Novaloc™ processing spacer* on the locator-type abutment
- fabrication of the prosthetic restoration

* Please follow the relevant instructions of the manufacturer Valoc (www.valoc.ch).



Left: Novaloc™ forming/fixing-matrix, right: Novaloc™ processing spacer
Manufacturer: Valoc (www.valoc.ch)



Left: Novaloc™ model analogue blue, right: Novaloc™ Model analogue angled 15°
Manufacturer: Valoc (www.valoc.ch)

Prosthetic aftercare of the Z5-BL implants

Regular prosthetic aftercare of Z5-BL implants is necessary as with all implant systems. As individual factors such as the patient's oral hygiene, cooperation, etc. play a major role in determining regular prosthetic aftercare, the interval proposed here can only be regarded as a guideline.

On the day of final placement of the restoration

- Repeat check for impression material residues
- Check the cement in the sulcus area
- Static and dynamic occlusion check
- Oral hygiene instruction
- X-ray examination

1 day after placement of the restoration

- Check the cement in the sulcus area
- Static and dynamic occlusion check
- Oral hygiene instruction

3 months after placement of the restoration

- Check for plaque
- Static and dynamic occlusion check
- Hygiene check; if necessary reinstruction and motivation
- Performance of a prophylaxis
- For removable prosthetic restorations, check resilience and perform relining if necessary

6 months after placement of the restoration

- Check for plaque
- Static and dynamic occlusion check
- Hygiene check; if necessary reinstruction and motivation
- X-ray examination
- Performance of a prophylaxis
- For removable prosthetic restorations, check resilience and perform relining if necessary

> Check-up every 6 months**> Regular prophylaxis**

Cementing of restorations on Z5-BL implants

General note

The following points must be observed when fixing temporary or final restorations on Z5-BL implants:

- Relative drainage of the working area
- Completely remove blood and/or saliva
- Cement residues must be completely removed
- Clean the peri-implant sulcus completely of cement residues (probe, superfloss)
- Temporary cementation of final bridge constructions carries the risk of a one-sided loosening of a bridge anchor with an increased risk of a possible fracture of the bridge or abutment ceramic.

Final cementing on Z5-BL implants

Z-SYSTEMS recommends the use of cements for final cementation that are suitable for zirconium oxide cementation. Zirconium oxide cannot be roughened intraorally by known adhesive systems.

NOTE: The temporary cementing of final restorations is not recommended.

Z-SYSTEMS accepts no liability for incorrect use of fastening systems or damage to the prosthetic restoration and/or to the implant itself resulting therefrom.

Prophylaxis for Z5-BL implants

Zirconium oxide has a very low affinity for plaque. Therefore, compared to other materials used in dentistry, there is very little plaque on Z5-BL implants. Nevertheless, regular and adequate prophylaxis is also indispensable for Z5-BL implants.

Due to their special material and design, some points deviating from the usual prophylaxis guidelines for implants must be observed with Z5-BL implants.

NOTE: Use only Teflon-based hand scalers and curettes for cleaning Z5-BL implants.

Rinsing solutions based on chlorhexidine and/or alcohol can be used in the short-term without concern. These solutions are not recommended for long-term use due to possible discolouration of the tooth structure and cement gaps.


Do not use ultrasound-operated, metallic cleaning aids to clean Z5-BL implants. Always avoid the application of ultrasound to Z5-BL implants through metallic carriers. Improper use and application of ultrasound can cause lasting damage to the surface of the Z5-BL implant.

When working with metallic cleaning aids (ultrasound-operated scalers or hand-curettes or scalers) there is the possibility of metallic abrasion on the implant surface. This abrasion is difficult or impossible to remove.

Do not use abrasive prophylaxis pastes to clean Z5-BL implants. A powder/water jet cleaner (Air-Flow®) is not suitable for cleaning Z5-BL implants.

How to remove a ceramic screw fragment from a ceramic implant


If an abutment failure occurs or the instructions for screwing in the ceramic screw are ignored, the ceramic screw (BL-OSC-H) may break. After a fracture a screw fragment ❶ may remain in the internal thread of the Z5-BL implant. Possible fracture situations are shown in the following figure.




❶ screw fragment (BL-OSC-H)

Instructions for removing a ceramic screw fragment

Clamping the Diamond Grinder Round in the contra angle Handpiece



RECOMMENDATION:
Diamond Grinder "round", Medium Grit
1) Ø 1.4 ± 0.1 mm
2) L = 25 to 30 mm




Program setting of the drive unit:


Speed (rpm)	15-20
Direction of rotation	(M) left-handed counterclockwise

NOTE: For the safe turning out a new grade diamond grinder medium grain must be used.

The Diamond Grinder Round is pressed against the screw fragment (red) to unscrew the fragment at 15 rpm, as shown in the figure below with the steps from 1 to 5.



RECOMMENDATION:
Unscrew the screw fragment (red) by slightly eccentric guidance of the diamond grinder around the implant axis.



Notes

This image shows a single sheet of white paper with horizontal ruling lines. The lines are evenly spaced and run across the width of the page. There are no margins, text, or other markings on the paper.

Notes

[illegible]

Notes

This image shows a single sheet of white paper with horizontal ruling lines. The lines are evenly spaced and run across the width of the page. There are no margins, text, or other markings on the paper.



Headquarters

Z-Systems AG | Werkhofstrasse 5 | CH-4702 Oensingen | Tel. +41 62 388 69 69 | support@zsystems.com

US Distribution

Z-Systems USA Inc. | 284 Monponsett Street, Suite 209 | Halifax, MA 02338 | phone +1 (781) 754-6566 | contact@zsystems.com

www.zsystems.com